


SCHWINGGESCHWINDIGKEIT




DATEI: 2493-HS-5-11052026-203940_raw.csv

Aufzeichnungsdauer: 5.18 s | Abtastrate: 1000 Hz

Datum	11.05.2026
Zeit	22:39:40
Event-ID	2493-HS-5-11052026-203940
Status	Ereignis
X	1.08 mm/s @ 1.4 Hz
Y	0.58 mm/s @ 1.2 Hz
Z	2.94 mm/s @ 1.2 Hz
Vn	3.05 mm/s @ 1.2 Hz

SPEKTRALANALYSE - KOMPONENTEN FFT



Kanal	Spitzwert	Spektrum 25%	Spektrum 50%	Spektrum 75%
X	1.4 Hz	1.7 Hz	6.4 Hz	15.2 Hz
Y	1.2 Hz	1.4 Hz	8.3 Hz	28.4 Hz
Z	1.2 Hz	4.4 Hz	7.9 Hz	33.4 Hz

SPEKTRALANALYSE - FFT (DIN 4150-3)

